
# **CENTRAL** MONITORING STATION



# **FX 3000C**



www.emtel.pl

#### Central monitoring system allows collecting information from all patient monitors connected to the network together with FX 3000C.

Quick access to each patient monitor from a central station makes the work of the medical staff easier and more effective.

The system facilitates the communication between patient monitors and central station, remote changing of patient monitors' settings.

An intuitive software, operation via keyboard and mouse or touch screen make the use of the central station extremely simple. The central station enables print-outs from each patient monitor and reviewing large, full disclosure archives.

### **EUROPEAN PRODUCT** SOLD AROUND THE WORLD

Over 30 years of experience, thousands of patient monitors and defibrillators sold in Europe, Asia, Australia, Africa and South America.





# FX 3000C

## TECHNICAL PARAMETERS

#### GENERAL PARAMETERS

Power supply Internal battery Protection class Safety requirements 110/230 VAC ±10% 50/60 Hz 12 VDC I, CF MDD: 93/42/EEC, 2007/47/EC; EN 60601-1

#### LCD DISPLAY

Colour display Trends LCD TFT 19", 24" 24 - 120 h

#### **VIEW MODE**

Full view Selective view Selective single view all patients 4, 8 patients or other 1 patient (copy of bedside monitor screen)

#### TRANSMISSION TYPE

Ethernet (wireless option)

#### ADDITIONAL INFORMATION

Operation via keyboard and mouse Connection to laser printer (A-4 format) Real time printout and from archive Printout directly from central station or from patient monitor Full disclosure archive Audible and visible alarms Automatic identification of patient monitor where an alarm was activated Changing and adjusting alarms' settings of patient monitor from the central station

#### ADDITIONAL OPTIONS

Touch screen Cooperation with telemetry modules Additional second screen Alarms archive Transfer of patient data to PC Possibility of second and third monitoring level





FX 3000C brochure v.1.1.EN



Emtel reserves the right to make changes to this sheet and the product at any time without notice.





EMTEL Śliwa sp. k. A. Mickiewicza 66, 41-807 Zabrze, POLAND phone: +48 32 271 90 13, fax: +48 32 271 57 27 medical@emtel. pl



www.emtel.pl